

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	25.4	
2 - 1	24.9	
3 - 1	24	
4 - 1	24.5	
6 - 1	25	
7 - 1	24.4	
8 - 1	25.4	
9 - 1	25.2	
10 - 1	24.8	
11 - 1	24	
13 - 1	25	
16 - 1	25.4	
17 - 1	25	
18 - 1	24.7	
20 - 1	24	
21 - 1	25	
22 - 1	24.2	
27 - 1	25.3	
28 - 1	24	
29 - 1	24.3	
30 - 1	25.4	
31 - 1	24.5	

Westgard	تفسیر	V D3	آزمایش	HPLC Agilent 1100 -1	دستگاه	3	ماه	1402	سال
Control	آزمایشگاه	Mean	24.71	SD	0.77	Exp. Date	1402/3/31	Lot. No	
Next Month	Mean	24.75	SD	0.51	CV	2.06			

