

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 1.74	
Mean	9.49	Mean	14.25
SD	0.17	SD	0.25
CV	1.77	CV	1.74
Day-Run	Value	Day-Run	Value
1 - 1	9.3	1 - 1	14.6
2 - 1	9.4	2 - 1	14.5
3 - 1	9.5	3 - 1	14.5
6 - 1	9.2	6 - 1	14.2
8 - 1	9.6	8 - 1	14.5
9 - 1	9.7	9 - 1	14.3
10 - 1	9.4	10 - 1	14.1
11 - 1	9.7	11 - 1	13.9
17 - 1	9.2	17 - 1	14
20 - 1	9.6	20 - 1	14.3
21 - 1	9.6	21 - 1	14.1
22 - 1	9.3	22 - 1	14.1
23 - 1	9.5	23 - 1	14.5
24 - 1	9.6	24 - 1	13.6
25 - 1	9.3	25 - 1	14.4
27 - 1	9.7	27 - 1	14.2
28 - 1	9.6	28 - 1	14.1
29 - 1	9.4	29 - 1	14.3
30 - 1	9.5	30 - 1	14.3
31 - 1	9.7	31 - 1	14.5

Westgard	تفسیر	Calcium	آزمایش	اتوآنالایزر BS-800	دستگاه	3	ماه	1402	سال
Normal Control	biorex	Mean	9.52	SD	0.13	Exp. Date	1403/1/17	Lot. No	035

