

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	43.2	
2 - 1	44	
3 - 1	43.1	
4 - 1	42.9	
7 - 1	42.8	
8 - 1	42.1	
9 - 1	43	
10 - 1	42.1	
11 - 2	42.4	
12 - 3	43.1	
14 - 1	42.1	
15 - 1	43	
16 - 1	44	
17 - 1	44.1	
18 - 2	44.1	
19 - 3	43.2	
21 - 1	43.5	
22 - 1	44	
23 - 1	43	
28 - 1	43.9	
29 - 1	44.1	

Westgard	تفسیر	Lactate(Lactic Acid)	آزمایش	BS-480 اتوآنالایزر	دستگاه	5	ماه	1402	سال
Control	biorex	Mean	43.09	SD	0.66	Exp. Date	1403/3/1	Lot. No	
Next Month	Mean	43.22	SD	0.69	CV	1.6			

