

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	222.39
SD	4.65
CV	2.09
Day-Run	Value
1 - 1	218
3 - 1	219
4 - 1	221
5 - 1	227
6 - 1	223
7 - 1	235
10 - 1	219
11 - 1	231
12 - 1	226
13 - 1	224
14 - 1	222
15 - 1	215
16 - 1	225
17 - 1	221
18 - 1	223
20 - 1	219
21 - 1	221
22 - 1	220
24 - 1	228
25 - 1	219
26 - 1	222
28 - 1	217
29 - 1	220

High - CV : 3.36	
Mean	550.05
SD	18.47
CV	3.36
Day-Run	Value
1 - 1	555
3 - 1	561
4 - 1	554
5 - 1	566
6 - 1	548
7 - 1	543
10 - 1	548
11 - 1	566
12 - 1	547
13 - 1	594
14 - 1	534
15 - 1	550
16 - 1	559
17 - 1	538
18 - 1	559
21 - 1	528
22 - 1	544
24 - 1	522
25 - 1	515
26 - 1	530
28 - 1	560
29 - 1	580

Westgard	تفسیر	PLT	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	4	ماه	1402	سال
Normal Control		Mean	224.3	SD	5.56	Exp. Date	1402/4/13	Lot. No	

