

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 1.25	
Mean	176.28	Mean	532.97
SD	2.6	SD	6.64
CV	1.47	CV	1.25
Day-Run	Value	Day-Run	Value
3 - 1	177.2	3 - 1	544.2
4 - 1	172.7	4 - 1	531.5
5 - 1	177.1	5 - 1	541
6 - 1	176.4	6 - 1	527.9
8 - 1	175.9	8 - 1	528.9
9 - 1	180	9 - 1	542.6
10 - 1	176.9	10 - 1	533.4
12 - 1	176.7	12 - 1	535.3
13 - 1	172.1	13 - 1	523
15 - 1	177	15 - 1	538.1
16 - 1	175.7	16 - 1	531
17 - 1	174.8	17 - 1	535.6
20 - 1	170	20 - 1	530
22 - 1	180	22 - 1	531.6
23 - 1	179.2	23 - 1	529.5
24 - 1	178	24 - 1	540
25 - 1	177	25 - 1	528
26 - 1	175	26 - 1	538.2
27 - 1	174.1	27 - 1	531.5
29 - 1	180	29 - 1	534.8
30 - 1	176.1	30 - 1	516.2

Westgard	تفسیر	Ikalan Phosphatase(ALP)	آزمایش	اتوانالایزر BS-800	دستگاه	7	ماه	1402	سال
Normal Control	biorex	Mean	175.4	SD	2.4	Exp. Date	1403/1/17	Lot. No	035

