

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.82	
Mean	173.4	Mean	539.5
SD	2.4	SD	9.83
CV	1.39	CV	1.82
Day-Run	Value	Day-Run	Value
1 - 1	173.9	1 - 1	542.9
2 - 1	170	2 - 1	521.8
4 - 1	175.4	4 - 1	560
5 - 1	176.4	5 - 1	545
6 - 1	173.8	6 - 1	539
7 - 1	173.4	7 - 1	531.2
8 - 1	175.8	8 - 1	541.8
11 - 1	170	11 - 1	559
12 - 1	172.4	12 - 1	531
13 - 1	172.4	13 - 1	523.5
14 - 1	175.3	14 - 1	547.9
18 - 1	174.2	18 - 1	545.8
19 - 1	173.3	19 - 1	541.6
20 - 1	171.3	20 - 1	538
21 - 1	170.4	21 - 1	545.1
22 - 1	177	22 - 1	532.2
26 - 1	170	26 - 1	538
27 - 1	178	27 - 1	531.7
28 - 1	172.9	28 - 1	535.9
29 - 1	172	29 - 1	538.6

Westgard	تفسیر	Ikalan Phosphatase(ALP)	آزمایش	اتوآنالایزر BS-800	دستگاه	6	ماه	1402	سال
Normal Control	biorex	Mean	175	SD	3.29	Exp. Date	1403/1/17	Lot. No	035

