

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 1.21	
Mean	38.73	Mean	143.82
SD	0.81	SD	1.74
CV	2.08	CV	1.21
Day-Run	Value	Day-Run	Value
1 - 1	38.4	1 - 1	140.2
2 - 1	38.9	2 - 1	145.2
4 - 1	37.4	4 - 1	142.6
5 - 1	36.8	5 - 1	144.2
6 - 1	39	6 - 1	143
7 - 1	38.2	7 - 1	145.2
8 - 1	39.4	8 - 1	143
11 - 1	39.9	11 - 1	147
12 - 1	39.4	12 - 1	145
13 - 1	39.9	13 - 1	142.1
14 - 1	39.4	14 - 1	143.8
18 - 1	37.8	18 - 1	143.3
19 - 1	38.5	19 - 1	145.2
20 - 1	39	20 - 1	144
21 - 1	38.8	21 - 1	143.3
22 - 1	38	22 - 1	147.3
26 - 1	39.2	26 - 1	142.2
27 - 1	39.4	27 - 1	145
28 - 1	38.5	28 - 1	142.8
29 - 1	38.6	29 - 1	142

Westgard	تفسیر	SGOT ( AST )	آزمایش	اتوآنالایزر BS-800	دستگاه	6	ماه	1402	سال
Normal Control	biorex	Mean	38.81	SD	0.91	Exp. Date	1403/1/17	Lot. No	035

