

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 1.76	
Mean	4.21	Mean	5.13
SD	0.05	SD	0.09
CV	1.28	CV	1.76
Day-Run	Value	Day-Run	Value
3 - 1	4.2	3 - 1	5.1
4 - 1	4.1	4 - 1	5
5 - 1	4.1	5 - 1	5.1
6 - 1	4.2	6 - 1	5.1
8 - 1	4.2	8 - 1	5.1
9 - 1	4.3	9 - 1	5.1
10 - 1	4.2	10 - 1	5.2
12 - 1	4.3	12 - 1	5.1
13 - 1	4.2	13 - 1	5
15 - 1	4.2	15 - 1	5.2
16 - 1	4.2	16 - 1	5.2
17 - 1	4.2	17 - 1	5.2
20 - 1	4.3	20 - 1	5
22 - 1	4.2	22 - 1	5.2
23 - 1	4.2	23 - 1	5.2
24 - 1	4.2	24 - 1	5
25 - 1	4.2	25 - 1	5.2
26 - 1	4.2	26 - 1	5.2
27 - 1	4.2	27 - 1	5.2
29 - 1	4.3	29 - 1	5
30 - 1	4.2	30 - 1	5.3

Westgard	تفسیر	Albomin	آزمایش	BS-800 اتوآنالایزر	دستگاه	7	ماه	1402	سال
Normal Control	biorex	Mean	4.2	SD	0.1	Exp. Date	1403/1/17	Lot. No	035

