

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
3 - 1	79.2	
4 - 1	83.8	
5 - 1	80	
6 - 1	78.6	
8 - 1	80	
9 - 1	77.8	
10 - 1	80.3	
10 - 1	80.3	
12 - 1	81.3	
13 - 1	79.4	
15 - 1	78.6	
16 - 1	81	
17 - 1	84	1:2S
20 - 1	81	
22 - 1	77.5	
23 - 1	78.7	
24 - 1	81	
25 - 1	77.3	
26 - 1	77	
27 - 1	80.2	
29 - 1	82.4	
30 - 1	84	1:2S

Westgard	تفسیر	Amylase	آزمایش	اتوآنالایزر BS-800	دستگاه	7	ماه	1402	سال
Control	biorex	Mean	79.97	SD	2.01	Exp. Date	1403/1/17	Lot. No	035
Next Month	Mean	80.15	SD	2.08	CV	2.59			

