

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 2.59	
Mean	1	Mean	4.27
SD	0.07	SD	0.11
CV	7.19	CV	2.59
Day-Run	Value	Day-Run	Value
3 - 1	1.05	3 - 1	4.25
4 - 1	1.1	4 - 1	4.37
5 - 1	1	5 - 1	4.31
6 - 1	0.92	6 - 1	4.07
8 - 1	1.02	8 - 1	4.36
9 - 1	1.1	9 - 1	4.25
10 - 1	0.9	10 - 1	4.17
12 - 1	1.07	12 - 1	4.4
13 - 1	0.91	13 - 1	4.07
15 - 1	1.03	15 - 1	4.28
16 - 1	0.93	16 - 1	4.3
17 - 1	1	17 - 1	4.15
20 - 1	1.1	20 - 1	4.38
22 - 1	1	22 - 1	4.31
23 - 1	1.12	23 - 1	4.18
24 - 1	0.9	24 - 1	4.5
25 - 1	1.07	25 - 1	4.16
26 - 1	1	26 - 1	4.3
27 - 1	0.93	27 - 1	4.16
29 - 1	0.95	29 - 1	4.3
30 - 1	1	30 - 1	4.3

Westgard	تفسیر	Bili T	آزمایش	اتوآنالایزر BS-800	دستگاه	7	ماه	1402	سال
Normal Control	biorex	Mean	0.99	SD	0.07	Exp. Date	1403/1/17	Lot. No	035

