

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	90.8	
3 - 1	90.7	
4 - 1	90.2	
5 - 1	90.4	
6 - 1	91.3	
8 - 1	90.3	
9 - 1	91.3	1:3S
10 - 1	91.4	
12 - 1	91.7	
13 - 1	91.4	
15 - 1	90.6	
16 - 1	90.5	
17 - 1	90.3	
18 - 1	91.1	
19 - 1	90.8	
20 - 1	91.3	
22 - 1	90.5	
23 - 1	90.6	
24 - 1	91.5	
25 - 1	91.1	
26 - 1	91.3	
27 - 1	91.8	1:2S
30 - 1	90.4	

Westgard	تفسیر	C3	آزمایش	BS-480 اتوآنالایزر	دستگاه	7	ماه	1402	سال
Control	biorex	Mean	90.9	SD	0.43	Exp. Date	1402/9/14	Lot. No	641-01-030/2
Next Month	Mean	90.93	SD	0.49	CV	0.54			

