

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	17.9	
3 - 1	17.8	
4 - 1	17.9	
5 - 1	18	
6 - 1	18.02	
8 - 1	18.06	
9 - 1	17.95	
10 - 1	18.09	
12 - 1	17.8	
13 - 1	17.9	
15 - 1	18.1	
16 - 1	18	
17 - 1	17.95	
18 - 1	18.04	
19 - 1	18.09	
20 - 1	17.9	
22 - 1	17.8	
23 - 1	17.9	
24 - 1	17.7	
25 - 1	17.5	1:2S
26 - 1	17.8	
27 - 1	17.9	
29 - 1	17.99	
30 - 1	18.06	

Westgard	تفسیر	C4	آزمایش	BS-480 اتوآنالایزر	دستگاه	7	ماه	1402	سال
Control	biorex	Mean	17.88	SD	0.13	Exp. Date	1402/9/14	Lot. No	641-01-030/2
Next Month	Mean	17.92	SD	0.14	CV	0.79			

