

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	51.4	
2 - 1	52.1	
4 - 1	50.8	
5 - 1	52.1	
6 - 1	48.5	
7 - 1	51.1	
8 - 1	52.75	
9 - 1	51.95	
11 - 1	51.74	
12 - 1	51.32	
13 - 1	48.5	
14 - 1	47.1	
16 - 1	50.7	
18 - 1	50.95	
19 - 1	48.9	
20 - 1	49.2	
21 - 1	48	
22 - 1	49.5	
26 - 1	48.7	
27 - 1	52.4	
28 - 1	52.1	
29 - 1	50.4	
30 - 1	51.4	

Westgard	تفسیر	CA 125	آزمایش	2000 ایمولایت XPI-	دستگاه	6	ماه	1402	سال
Control	Seronorm	Mean	49.97	SD	1.58	Exp. Date	1402/4/25	Lot. No	2011909
Next Month	Mean	50.5	SD	1.61	CV	3.2			

