

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 1.55	
Mean	127.68	Mean	215.5
SD	1.89	SD	3.34
CV	1.48	CV	1.55
Day-Run	Value	Day-Run	Value
3 - 1	126	3 - 1	210
4 - 1	130	4 - 1	216
5 - 1	128	5 - 1	210
6 - 1	130	6 - 1	213
8 - 1	125	8 - 1	215
9 - 1	123	9 - 1	213
10 - 1	128	10 - 1	212
12 - 1	126	12 - 1	218
13 - 1	127	13 - 1	220
15 - 1	126	15 - 1	211
16 - 1	129	16 - 1	217
17 - 1	127	17 - 1	216
20 - 1	129.2	20 - 1	213.5
22 - 1	130	22 - 1	218
23 - 1	128	23 - 1	220
24 - 1	128	24 - 1	219
25 - 1	127	25 - 1	220
26 - 1	131	26 - 1	216
27 - 1	128	27 - 1	214
29 - 1	128	29 - 1	214
30 - 1	127	30 - 1	220

Westgard	تفسیر	CPK	آزمایش	اتوانالایزر BS-800	دستگاه	7	ماه	1402	سال
Normal Control	biorex	Mean	127.35	SD	2.16	Exp. Date	1403/1/17	Lot. No	035

