

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
3 - 1	127	
4 - 1	126	
5 - 1	126	
6 - 1	125	
8 - 1	128	
9 - 1	126	
10 - 1	127	
12 - 1	126	
13 - 1	128	
15 - 1	127	
16 - 1	125	
17 - 1	124	
20 - 1	130	1:2S
22 - 1	126	
23 - 1	130	1:2S
24 - 1	124	
25 - 1	127	
26 - 1	127	
27 - 1	127	
29 - 1	127	
30 - 1	128	

Westgard	تفسیر	Copper (Cu)	آزمایش	اتوآنالایزر BS-800	دستگاه	7	ماه	1402	سال
Control	biorex	Mean	126.25	SD	1.61	Exp. Date	1402/9/20	Lot. No	hsn.35
Next Month	Mean	126.71	SD	1.59	CV	1.25			

