

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	24.5	
2 - 1	24.9	
4 - 1	25.7	
5 - 1	25.9	
6 - 1	25.4	
7 - 1	25.1	
8 - 1	25.94	
9 - 1	26.2	
11 - 1	25.4	
12 - 1	25.51	
13 - 1	23.2	
14 - 1	22.9	
16 - 1	25	
18 - 1	25.7	
19 - 1	25.5	
20 - 1	24.8	
21 - 1	24	
22 - 1	26.2	
26 - 1	25.01	
27 - 1	26.4	
28 - 1	26.1	
29 - 1	24.9	
30 - 1	25	

Westgard	تفسیر	FSH	آزمایش	2000 ایمولایت XPI-	دستگاه	6	ماه	1402	سال
Control	Seronorm	Mean	25.42	SD	1.33	Exp. Date	1402/6/5	Lot. No	2011909
Next Month	Mean	25.19	SD	0.9	CV	3.56			

