

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	1.54	
3 - 1	1.45	
4 - 1	1.52	
5 - 1	1.32	
6 - 1	1.39	
7 - 1	1.52	
8 - 1	1.48	
9 - 1	1.37	
10 - 1	1.39	
12 - 1	1.45	
13 - 1	1.48	
15 - 1	1.52	
16 - 1	1.42	
17 - 1	1.54	
18 - 1	1.49	
19 - 1	1.45	
20 - 1	1.49	
22 - 1	1.35	
22 - 2	1.59	1:2S
22 - 3	1.45	
23 - 1	1.34	
24 - 1	1.39	
25 - 1	1.37	
26 - 1	1.41	
27 - 1	1.52	
29 - 1	1.56	
30 - 1	1.55	

Westgard	تفسیر	Free PSA	آزمایش	2000 ایمولایت XPI-	دستگاه	7	ماه	1402	سال
Control	biorex	Mean	1.44	SD	0.06	Exp. Date	1402/7/1	Lot. No	365-02-028
Next Month	Mean	1.46	SD	0.07	CV	5.14			

