

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

| Day-Run | Value | Description |
|---------|-------|-------------|
| 2 - 1 | 70.1 | |
| 4 - 1 | 66.4 | |
| 5 - 1 | 69.7 | |
| 6 - 1 | 66.4 | |
| 7 - 1 | 67.8 | |
| 8 - 1 | 71.1 | |
| 9 - 1 | 67.5 | |
| 11 - 1 | 69 | |
| 12 - 1 | 67.1 | |
| 13 - 1 | 66.4 | |
| 14 - 1 | 67 | |
| 18 - 1 | 69 | |
| 19 - 1 | 70.5 | 1:3S |
| 20 - 1 | 70 | |
| 21 - 1 | 67.8 | |
| 22 - 1 | 70 | |
| 26 - 1 | 67.5 | |
| 27 - 1 | 70.5 | |
| 28 - 1 | 70 | |
| 29 - 1 | 69.4 | |

| | | | | | | | | | |
|------------|--------|----------|--------|--------------------|--------|-----------|-----------|---------|-----|
| Westgard | تفسیر | Gamma GT | آزمایش | اتوآنالایزر BS-800 | دستگاه | 6 | ماه | 1402 | سال |
| Control | biorex | Mean | 68.67 | SD | 1.36 | Exp. Date | 1403/1/17 | Lot. No | 035 |
| Next Month | Mean | 68.66 | SD | 1.57 | CV | 2.28 | | | |

