

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal	
Mean	12.16
SD	0.15
CV	1.23
Day-Run	Value
1 - 1	12.2
2 - 1	11.9
4 - 1	12.3
5 - 1	12.2
6 - 1	12.2
7 - 1	12
8 - 1	12.3
9 - 1	12.3
11 - 1	11.9
12 - 1	12
13 - 1	12.2
14 - 1	12.3
18 - 1	12.4
19 - 1	12.3
20 - 1	12
21 - 1	12
22 - 1	12.2
28 - 1	12.3
30 - 1	12.1
27 - 1	12
28 - 2	12.1
29 - 1	12.3

High - CV : 1.23	
Mean	17.55
SD	0.22
CV	1.23
Day-Run	Value
1 - 1	17.5
2 - 1	17.6
4 - 1	17.7
5 - 1	17.3
6 - 1	17.5
7 - 1	17.4
8 - 1	17.5
9 - 1	17.6
11 - 1	17.7
12 - 1	17.9
13 - 1	17.3
14 - 1	17.7
18 - 1	17.8
20 - 1	18
21 - 1	17.3
22 - 1	17.7
28 - 1	17.2
30 - 1	17.7
27 - 1	17.6
28 - 2	17.2
29 - 1	17.5

Westgard	تفسیر	HB	آزمایش	NIHON KOHDEN سل کانتیر	دستگاه	6	ماه	1402	سال
Normal Control	Labex	Mean	12.14	SD	0.15	Exp. Date	1402/6/30	Lot. No	02-2305n

