

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	132	
4 - 1	132.8	
5 - 1	138.3	
6 - 1	139	
7 - 1	140.4	
8 - 1	138.5	4:1S
9 - 1	133.5	
11 - 1	135.4	
12 - 1	137.4	
13 - 1	136.1	
14 - 1	131	
14 - 2	130.8	
16 - 1	135.4	
18 - 1	139	
19 - 1	132.8	
20 - 1	137	
21 - 1	140	
22 - 1	138.7	
26 - 1	137.7	
27 - 1	135.4	
29 - 1	135.9	
30 - 1	130.2	

Westgard	تفسیر	Ig A	آزمایش	BS-480 اتوآنالایزر	دستگاه	6	ماه	1402	سال
Control	biorex	Mean	135.33	SD	2.67	Exp. Date	1403/9/14	Lot. No	641-01-030/2
Next Month	Mean	135.79	SD	3.14	CV	2.31			

