

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	757	
4 - 1	755	
5 - 1	776	Cusum 3S
6 - 1	760	1:2S
7 - 1	758	
8 - 1	752	
9 - 1	726	
11 - 1	728	
12 - 1	786	
13 - 1	724	
14 - 1	767	
16 - 1	795	
18 - 1	763	
19 - 1	688.9	
20 - 1	779	
21 - 1	786	
22 - 1	767	
26 - 1	723	
27 - 1	780.5	
28 - 1	705.6	
29 - 1	734	
30 - 1	744.2	

Westgard	تفسیر	Ig G	آزمایش	BS-480 اتوآنالایزر	دستگاه	6	ماه	1402	سال
Control	biorex	Mean	725.77	SD	34.7	Exp. Date	1403/2/17	Lot. No	641-01-030/2
Next Month	Mean	752.51	SD	27.92	CV	3.71			

