

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	58.3	
2 - 1	59.6	
4 - 1	65.5	
5 - 1	70.2	
7 - 1	67	
8 - 1	61.7	
9 - 1	58.3	
11 - 1	60.2	
12 - 1	65.1	
13 - 1	68.8	
14 - 1	57.9	
16 - 1	72.2	
18 - 1	73.6	
19 - 1	64.9	
20 - 1	59.5	
21 - 1	55.3	
22 - 1	61.3	
26 - 1	63.2	
27 - 1	65.7	
28 - 1	59.2	
29 - 1	60.7	
30 - 1	62.9	

Westgard	تفسیر	Insulin (Base)	آزمایش	SIEMENS Centaur CP	دستگاه	6	ماه	1402	سال
Control	biorex	Mean	62.15	SD	6.37	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	63.23	SD	4.9	CV	7.75			

