

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
3 - 1	225.9	
4 - 1	235.3	
5 - 1	232.9	
6 - 1	234.5	
8 - 1	234.1	
9 - 1	240	
10 - 1	236.4	
12 - 1	230.8	
13 - 1	236	
15 - 1	239	
16 - 1	240	
17 - 1	231	
20 - 1	233	
22 - 1	230	
23 - 1	234.3	
24 - 1	235	
25 - 1	235.1	
26 - 1	236	
27 - 1	226	
29 - 1	235.9	
30 - 1	233	

Westgard	تفسیر	Iron	آزمایش	BS-800 اتوآنالایزر	دستگاه	7	ماه	1402	سال
Control	biorex	Mean	234.15	SD	4.79	Exp. Date	1403/1/17	Lot. No	035
Next Month	Mean	234.01	SD	3.79	CV	1.62			

