

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

| Day-Run | Value | Description |
|---------|-------|-------------|
| 1 - 1   | 32.5  |             |
| 2 - 1   | 34.1  |             |
| 4 - 1   | 32.9  |             |
| 5 - 1   | 32.1  |             |
| 6 - 1   | 31.7  |             |
| 7 - 1   | 31.9  |             |
| 8 - 1   | 33.85 |             |
| 9 - 1   | 33.65 |             |
| 11 - 1  | 32.45 |             |
| 12 - 1  | 31.95 |             |
| 13 - 1  | 31.6  |             |
| 14 - 1  | 35    |             |
| 16 - 1  | 34.7  |             |
| 18 - 1  | 34.5  |             |
| 19 - 1  | 31.4  |             |
| 20 - 1  | 30.9  |             |
| 21 - 1  | 30.54 |             |
| 22 - 1  | 31.6  |             |
| 26 - 1  | 31.4  |             |
| 27 - 1  | 32.7  |             |
| 28 - 1  | 32.6  |             |
| 29 - 1  | 34.2  |             |
| 30 - 1  | 32.4  |             |

|            |          |       |        |                    |        |           |           |         |         |
|------------|----------|-------|--------|--------------------|--------|-----------|-----------|---------|---------|
| Westgard   | تفسیر    | LH    | آزمایش | 2000 ایمولایت XPI- | دستگاه | 6         | ماه       | 1402    | سال     |
| Control    | Seronorm | Mean  | 32.29  | SD                 | 1.36   | Exp. Date | 1403/5/31 | Lot. No | 2011909 |
| Next Month | Mean     | 32.64 | SD     | 1.27               | CV     | 3.89      |           |         |         |

