

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	67.6	
2 - 1	69	
4 - 1	65	
5 - 1	64.8	
6 - 1	70	
7 - 1	71	
8 - 1	69	
11 - 1	66.5	
12 - 1	67.8	
13 - 1	67	
14 - 1	69	
18 - 1	65	
19 - 1	64.9	
20 - 1	65.6	
21 - 1	69.1	
22 - 1	69.1	
26 - 1	68.2	
27 - 1	68.9	
28 - 1	66.8	
29 - 1	68	

Westgard	تفسیر	Lipase	آزمایش	BS-800 اتوآنالایزر	دستگاه	6	ماه	1402	سال
Control	biorex	Mean	67.35	SD	1.97	Exp. Date	1403/1/17	Lot. No	035
Next Month	Mean	67.62	SD	1.85	CV	2.73			

