

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal	
Mean	28.12
SD	0.3
CV	1.06
Day-Run	Value
1 - 1	28
2 - 1	28.5
4 - 1	28
5 - 1	27.9
6 - 1	28.5
7 - 1	28.2
8 - 1	28.5
9 - 1	28
11 - 1	27.7
12 - 1	28.3
13 - 1	28.2
14 - 1	28.5
18 - 1	28.1
20 - 1	28.5
20 - 2	28.2
21 - 1	28.3
22 - 1	27.9
28 - 1	28
26 - 1	27.8
27 - 1	27.7
28 - 2	28.4
29 - 1	27.5

High - CV : 0.74	
Mean	31.67
SD	0.23
CV	0.74
Day-Run	Value
1 - 1	32
2 - 1	31.5
4 - 1	31.2
5 - 1	31.5
6 - 1	31.9
7 - 1	31.5
8 - 1	31.5
9 - 1	32
11 - 1	31.6
12 - 1	32
13 - 1	31.6
18 - 1	31.8
20 - 1	31.8
20 - 2	31.5
21 - 1	32
22 - 1	31.5
28 - 1	31.6
26 - 1	31.5
27 - 1	31.5
28 - 2	32
29 - 1	31.6

Westgard	تفسیر	MCH	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	6	ماه	1402	سال
Normal Control	Labex	Mean	28.07	SD	0.36	Exp. Date	1402/6/30	Lot. No	02-2305n

