

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

| Normal  |       |
|---------|-------|
| Mean    | 30.32 |
| SD      | 0.65  |
| CV      | 2.15  |
| Day-Run | Value |
| 1 - 1   | 30    |
| 2 - 1   | 30.5  |
| 4 - 1   | 31    |
| 5 - 1   | 29.5  |
| 6 - 1   | 30    |
| 7 - 1   | 31    |
| 8 - 1   | 29.8  |
| 9 - 1   | 29.5  |
| 11 - 1  | 30    |
| 12 - 1  | 30.6  |
| 13 - 1  | 30.6  |
| 14 - 1  | 29.4  |
| 18 - 1  | 30    |
| 19 - 1  | 30    |
| 20 - 1  | 29.9  |
| 20 - 2  | 32    |
| 21 - 1  | 30.5  |
| 22 - 1  | 29.5  |
| 28 - 1  | 31    |
| 26 - 1  | 31    |
| 27 - 1  | 30    |
| 28 - 2  | 31    |
| 29 - 1  | 30.5  |

| High - CV : 0.75 |       |
|------------------|-------|
| Mean             | 31.83 |
| SD               | 0.24  |
| CV               | 0.75  |
| Day-Run          | Value |
| 1 - 1            | 22.5  |
| 1 - 2            | 31.5  |
| 2 - 1            | 32    |
| 4 - 1            | 32    |
| 5 - 1            | 31.6  |
| 6 - 1            | 31.9  |
| 7 - 1            | 31.5  |
| 8 - 1            | 32    |
| 9 - 1            | 32    |
| 11 - 1           | 32    |
| 12 - 1           | 31.8  |
| 13 - 1           | 32    |
| 14 - 1           | 31.5  |
| 18 - 1           | 32    |
| 19 - 1           | 32    |
| 21 - 1           | 32    |
| 22 - 1           | 31.5  |
| 28 - 1           | 31.5  |
| 26 - 1           | 32    |
| 27 - 1           | 32.2  |
| 28 - 2           | 31.9  |
| 29 - 1           | 31.5  |

|                |       |      |        |                       |        |           |           |         |          |
|----------------|-------|------|--------|-----------------------|--------|-----------|-----------|---------|----------|
| Westgard       | تفسیر | MCHC | آزمایش | NIHON KOHDEN سل کانتر | دستگاه | 6         | ماه       | 1402    | سال      |
| Normal Control | Labex | Mean | 30.49  | SD                    | 0.57   | Exp. Date | 1402/6/30 | Lot. No | 02-2305n |

