

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	2.22
SD	0.14
CV	6.26
Day-Run	Value
1 - 1	2.2
2 - 1	2.1
4 - 1	2.3
5 - 1	2.2
6 - 1	2.3
7 - 1	2.3
8 - 1	2.2
11 - 1	2.3
12 - 1	2
13 - 1	2.3
14 - 1	1.9
18 - 1	2.4
19 - 1	2.2
20 - 1	2
21 - 1	2.4
22 - 1	2.1
26 - 1	2.2
27 - 1	2.3
28 - 1	2.2
29 - 1	2.4

High - CV : 3.29	
Mean	3.92
SD	0.13
CV	3.29
Day-Run	Value
1 - 1	3.7
2 - 1	3.9
4 - 1	4
5 - 1	3.8
6 - 1	3.8
7 - 1	4
8 - 1	3.9
11 - 1	3.9
12 - 1	4.2
13 - 1	3.8
14 - 1	4.1
18 - 1	4
19 - 1	3.8
20 - 1	4
21 - 1	4
22 - 1	3.8
22 - 2	3.8
26 - 1	3.9
27 - 1	4
28 - 1	3.8
29 - 1	4.1

Westgard	تفسیر	Magnesium(Mg)	آزمایش	BS-800 اتوآنالایزر	دستگاه	6	ماه	1402	سال
Normal Control	biorex	Mean	2.2	SD	0.16	Exp. Date	1403/1/17	Lot. No	035

