

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 3.03	
Mean	3.47	Mean	8.1
SD	0.11	SD	0.25
CV	3.17	CV	3.03
Day-Run	Value	Day-Run	Value
3 - 1	3.5	3 - 1	8
4 - 1	3.4	4 - 1	8
5 - 1	3.5	5 - 1	8
6 - 1	3.6	6 - 1	8
8 - 1	3.4	8 - 1	8.1
9 - 1	3.6	9 - 1	8
10 - 1	3.6	10 - 1	8.1
12 - 1	3.6	12 - 1	7.9
13 - 1	3.4	13 - 1	8.2
15 - 1	3.5	15 - 1	7.8
16 - 1	3.5	16 - 1	8.1
17 - 1	3.7	17 - 1	7.6
20 - 1	3.4	20 - 1	8
22 - 1	3.3	22 - 1	8.4
23 - 1	3.6	23 - 1	8.5
24 - 1	3.4	24 - 1	8.5
25 - 1	3.4	25 - 1	7.9
26 - 1	3.4	26 - 1	8.2
27 - 1	3.4	27 - 1	8
29 - 1	3.4	29 - 1	8.3
30 - 1	3.3	30 - 1	8.6

Westgard	تفسیر	Phosphorus (P)	آزمایش	BS-800 اتوآنالایزر	دستگاه	7	ماه	1402	سال
Normal Control	biorex	Mean	3.51	SD	0.15	Exp. Date	1403/1/17	Lot. No	035

