

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	4.6	
2 - 1	4.7	1:3S Cusum 3S
4 - 1	4.8	
5 - 1	4.7	
6 - 1	4.8	
7 - 1	4.7	
8 - 1	4.8	
9 - 1	4.7	
11 - 2	4.6	
12 - 1	4.6	
13 - 1	4.6	
14 - 1	4.71	
16 - 1	4.6	
18 - 1	4.7	
19 - 1	4.6	
20 - 1	4.7	
21 - 1	4.6	
22 - 1	4.8	
26 - 1	4.8	
27 - 1	4.7	

Westgard	تفسیر	Potassium ( K )	آزمایش	AUDICOM	دستگاه	6	ماه	1402	سال
Control	زیست شیمی	Mean	4.7	SD	0.1	Exp. Date	1404/1/14	Lot. No	761LOT17
Next Month	Mean	4.69	SD	0.08	CV	1.68			

