

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	4.7	
3 - 1	4.8	
4 - 1	4.68	
5 - 1	4.7	
6 - 1	4.6	
8 - 1	4.6	
9 - 1	4.69	
10 - 1	4.72	
12 - 2	4.8	
13 - 1	4.73	
15 - 1	4.7	
16 - 1	4.65	
17 - 1	4.64	
18 - 1	4.72	
19 - 1	4.77	
20 - 1	4.7	
22 - 1	4.7	
23 - 1	4.75	
24 - 1	4.7	
25 - 1	4.68	
26 - 1	4.7	
27 - 1	4.8	
30 - 1	4.78	

Westgard	تفسیر	Potassium (K)	آزمایش	AUDICOM	دستگاه	7	ماه	1402	سال
Control	زیست شیمی	Mean	4.69	SD	0.08	Exp. Date	1404/1/14	Lot. No	761LOT17
Next Month	Mean	4.7	SD	0.06	CV	1.25			

