

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	10.4	
3 - 1	10.21	
4 - 1	10.7	
5 - 1	8.95	
6 - 1	9.52	
8 - 1	10.23	
9 - 1	10.23	
10 - 1	8.91	
12 - 1	9.11	
13 - 1	9.8	
15 - 1	10.3	
16 - 1	9.5	
17 - 1	10.2	
18 - 1	10	
19 - 1	9	
20 - 1	9.2	
22 - 1	8.9	
23 - 1	10.2	
24 - 1	10.2	
25 - 1	9.5	
26 - 1	10.6	
27 - 1	10.2	
29 - 1	10.7	

Westgard	تفسیر	Progesterone	آزمایش	SIEMENS Centaur CP	دستگاه	7	ماه	1402	سال
Control	biorex	Mean	9.94	SD	0.62	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	9.85	SD	0.61	CV	6.16			

