

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.97	
Mean	6.96	Mean	8.28
SD	0.07	SD	0.16
CV	1.07	CV	1.97
Day-Run	Value	Day-Run	Value
3 - 1	6.9	3 - 1	8.3
4 - 1	6.9	4 - 1	8.2
5 - 1	7	5 - 1	8.3
6 - 1	6.9	6 - 1	8
8 - 1	7	8 - 1	8.4
9 - 1	7.1	9 - 1	8
10 - 1	7	10 - 1	8.5
12 - 1	7.1	12 - 1	8
13 - 1	7	13 - 1	8.3
15 - 1	6.9	15 - 1	8.3
16 - 1	6.8	16 - 1	8.1
17 - 1	6.9	17 - 1	8.3
20 - 1	7	20 - 1	8.2
22 - 1	7	22 - 1	8.5
23 - 1	7	23 - 1	8.4
24 - 1	7	24 - 1	8.3
25 - 1	6.9	25 - 1	8.4
26 - 1	6.9	26 - 1	8.3
27 - 1	7	27 - 1	8.3
29 - 1	6.9	29 - 1	8.2
30 - 1	6.9	30 - 1	8.6

Westgard	تفسیر	Protein Total	آزمایش	BS-800 اتوآنالایزر	دستگاه	7	ماه	1402	سال
Normal Control	biorex	Mean	6.97	SD	0.1	Exp. Date	1403/1/17	Lot. No	035

