

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

| Normal  |       |
|---------|-------|
| Mean    | 6.97  |
| SD      | 0.1   |
| CV      | 1.49  |
| Day-Run | Value |
| 1 - 1   | 6.9   |
| 2 - 1   | 6.9   |
| 4 - 1   | 7     |
| 5 - 1   | 6.9   |
| 6 - 1   | 7.2   |
| 7 - 1   | 6.9   |
| 8 - 1   | 7     |
| 11 - 1  | 6.9   |
| 12 - 1  | 7.1   |
| 13 - 1  | 7     |
| 14 - 1  | 7     |
| 18 - 1  | 7.2   |
| 19 - 1  | 7     |
| 20 - 1  | 7     |
| 21 - 1  | 6.9   |
| 22 - 1  | 6.9   |
| 26 - 1  | 6.9   |
| 27 - 1  | 6.9   |
| 28 - 1  | 6.8   |
| 29 - 1  | 6.9   |

| High - CV : 1.63 |       |
|------------------|-------|
| Mean             | 8.07  |
| SD               | 0.13  |
| CV               | 1.63  |
| Day-Run          | Value |
| 1 - 1            | 8.2   |
| 2 - 1            | 8.1   |
| 4 - 1            | 8     |
| 5 - 1            | 8.3   |
| 6 - 1            | 8.1   |
| 7 - 1            | 8.1   |
| 7 - 2            | 8.1   |
| 8 - 1            | 8.2   |
| 12 - 1           | 8.2   |
| 12 - 2           | 7.9   |
| 13 - 1           | 7.8   |
| 14 - 1           | 8     |
| 18 - 1           | 8.2   |
| 19 - 1           | 8.1   |
| 20 - 1           | 8     |
| 21 - 1           | 8     |
| 22 - 1           | 8.2   |
| 26 - 1           | 8     |
| 27 - 1           | 7.8   |
| 28 - 1           | 8     |
| 29 - 1           | 8.1   |

|                |        |               |        |                    |        |           |           |         |     |
|----------------|--------|---------------|--------|--------------------|--------|-----------|-----------|---------|-----|
| Westgard       | تفسیر  | Protein Total | آزمایش | اتوآنالایزر BS-800 | دستگاه | 6         | ماه       | 1402    | سال |
| Normal Control | biorex | Mean          | 7      | SD                 | 0.15   | Exp. Date | 1403/1/17 | Lot. No | 035 |

