

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	4.36
SD	0.05
CV	1.15
Day-Run	Value
1 - 1	4.25
2 - 1	4.4
4 - 1	4.4
5 - 1	4.38
6 - 1	4.37
7 - 1	4.3
8 - 1	4.41
9 - 1	4.42
11 - 1	4.34
12 - 1	4.25
13 - 1	4.34
14 - 1	4.39
18 - 1	4.42
19 - 1	4.34
20 - 1	4.36
21 - 1	4.34
22 - 1	4.4
28 - 1	4.41
26 - 1	4.36
27 - 1	4.31
28 - 2	4.4
29 - 1	4.33

High - CV : 1.3	
Mean	5.49
SD	0.07
CV	1.3
Day-Run	Value
1 - 1	5.4
2 - 1	5.5
4 - 1	5.55
5 - 1	5.45
6 - 1	5.5
7 - 1	5.54
8 - 1	5.53
9 - 1	5.44
11 - 1	5.5
12 - 1	5.5
13 - 1	5.5
14 - 1	5.37
18 - 1	5.43
19 - 1	5.5
20 - 1	5.5
21 - 1	5.6
22 - 1	5.64
28 - 1	5.42
26 - 1	5.55
27 - 1	5.4
28 - 2	5.42
29 - 1	5.6

Westgard	تفسیر	RBC	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	6	ماه	1402	سال
Normal Control	Labex	Mean	4.35	SD	0.05	Exp. Date	1402/6/30	Lot. No	02-2305n

