

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	37.6	
3 - 1	38	
4 - 1	36.3	
5 - 1	36.2	
7 - 1	39.1	
9 - 1	38.4	
10 - 1	35.2	
12 - 1	35.4	
13 - 1	39.11	
15 - 1	36	
16 - 1	34.3	
17 - 1	37.5	
18 - 1	35.5	
19 - 1	36.8	
20 - 1	36.9	
22 - 1	36.6	
23 - 1	35.6	
24 - 1	35	
25 - 1	37.9	
26 - 1	35.9	
27 - 1	38.6	
29 - 1	39.5	1:2S
30 - 1	38.8	

Westgard	تفسیر	RF	آزمایش	BS-480 اتوآنالایزر	دستگاه	7	ماه	1402	سال
Control	aptec	Mean	36.53	SD	1.36	Exp. Date	1404/7/11	Lot. No	05378
Next Month	Mean	36.97	SD	1.51	CV	4.08			

