

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	147.2	
3 - 1	147.5	
4 - 1	148.1	1:2S
5 - 1	146.9	
6 - 1	146.1	
8 - 1	146.6	
9 - 1	148.1	1:2S
10 - 1	146.3	
12 - 1	147.6	
13 - 1	146.5	
15 - 1	148	
16 - 1	147.3	
17 - 1	147.6	
18 - 1	146.6	
19 - 1	147.2	
20 - 1	147.2	
22 - 1	146	
23 - 1	146.5	
24 - 1	147.1	
25 - 1	147	
26 - 1	147.5	
27 - 1	147.4	
29 - 1	147.8	
28 - 1	147.2	

Westgard	تفسیر	Sodium (Na)	آزمایش	AUDICOM	دستگاه	7	ماه	1402	سال
Control	زیست شیمی	Mean	146.56	SD	0.75	Exp. Date	1404/1/14	Lot. No	761LOT17
Next Month	Mean	147.14	SD	0.6	CV	0.41			

