

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	2.37	
3 - 1	2.5	
4 - 1	2.48	
5 - 1	2.65	
6 - 1	2.25	
8 - 1	2.67	
9 - 1	2.53	
10 - 1	2.49	
12 - 1	2.25	
13 - 1	2.46	
15 - 1	2.65	
16 - 1	2.57	
17 - 1	2.66	
18 - 1	2.5	
19 - 1	2.65	
20 - 1	2.35	
22 - 1	2.76	
23 - 1	2.61	
24 - 1	2.38	
25 - 1	2.41	
26 - 1	2.32	
27 - 1	2.78	
29 - 1	2.61	

Westgard	تفسیر	T3	آزمایش	SIEMENS Centaur CP	دستگاه	7	ماه	1402	سال
Control	biorex	Mean	2.51	SD	0.16	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	2.52	SD	0.15	CV	6.06			

