

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	2.35	
2 - 1	2.73	
4 - 1	2.54	
5 - 1	2.43	
7 - 1	2.58	
8 - 1	2.36	
9 - 1	2.75	
11 - 1	2.7	Cusum 3S
12 - 1	2.45	
13 - 1	2.36	
14 - 1	2.75	
16 - 1	2.59	
18 - 1	2.5	
19 - 1	2.36	
20 - 1	2.22	
21 - 1	2.53	
22 - 1	2.65	
26 - 1	2.35	
27 - 1	2.49	
28 - 1	2.65	
29 - 1	2.57	
30 - 1	2.3	

Westgard	تفسیر	T3	آزمایش	SIEMENS Centaur CP	دستگاه	6	ماه	1402	سال
Control	biorex	Mean	2.49	SD	0.15	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	2.51	SD	0.16	CV	6.23			

