

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	11.7	
3 - 1	12	
4 - 1	10.7	
5 - 1	11.9	
6 - 1	11.4	Cusum 3S
8 - 1	11.1	
9 - 1	11.5	
10 - 1	12	
12 - 1	11.9	
13 - 1	11.6	
15 - 1	11	
16 - 1	10.5	
17 - 1	11.1	
18 - 1	11.5	
19 - 1	11.9	
20 - 1	11.2	
22 - 1	10.6	
23 - 1	11.6	
24 - 1	10.5	
25 - 1	10.8	
26 - 1	10.7	
27 - 1	11.6	
29 - 1	10.3	

Westgard	تفسیر	T4	آزمایش	SIEMENS Centaur CP	دستگاه	7	ماه	1402	سال
Control	biorex	Mean	11.2	SD	0.51	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	11.27	SD	0.54	CV	4.81			

