

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	3.76	
2 - 1	3.3	
4 - 1	3.89	
5 - 1	3.61	
7 - 1	3.51	
8 - 1	3.5	
9 - 1	3.77	
11 - 1	3.85	Cusum 3S
12 - 1	3.33	
13 - 1	3.85	
14 - 1	3.75	
16 - 1	3.79	
18 - 1	3.66	
19 - 1	3.34	
20 - 1	3.81	
21 - 1	3.86	
22 - 2	3.48	
22 - 1	3.86	4:1S
26 - 1	3.82	
27 - 1	3.6	
28 - 1	3.91	
29 - 1	3.39	
30 - 1	3.85	

Westgard	تفسیر	TSH	آزمایش	SIEMENS Centaur CP	دستگاه	6	ماه	1402	سال
Control	biorex	Mean	3.58	SD	0.18	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	3.67	SD	0.2	CV	5.49			

