

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.34	
Mean	34.24	Mean	131.1
SD	1.16	SD	1.75
CV	3.39	CV	1.34
Day-Run	Value	Day-Run	Value
3 - 1	32.4	3 - 1	130.5
4 - 1	33	4 - 1	131.8
5 - 1	32.4	5 - 1	133.3
6 - 1	36	6 - 1	133.7
8 - 1	35	8 - 1	130
9 - 1	34.1	9 - 1	133.1
10 - 1	34.2	10 - 1	128.1
12 - 1	34.6	12 - 1	130.7
13 - 1	35	13 - 1	128
15 - 1	34.7	15 - 1	133
16 - 1	32.2	16 - 1	130.1
17 - 1	36	17 - 1	130.1
20 - 1	35.2	20 - 1	130.2
22 - 1	32.5	22 - 1	133.2
23 - 1	34.5	23 - 1	131.6
24 - 1	33.7	24 - 1	130
25 - 1	35	25 - 1	131.8
26 - 1	34.6	26 - 1	130.7
27 - 1	35.2	27 - 1	131.4
29 - 1	33.8	29 - 1	128.5
30 - 1	35	30 - 1	133.4

Westgard	تفسیر	Urea	آزمایش	BS-800 اتوآنالایزر	دستگاه	7	ماه	1402	سال
Normal Control	biorex	Mean	34.36	SD	1.05	Exp. Date	1403/1/17	Lot. No	035

