

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 2	28.5	
2 - 1	29.3	
4 - 1	29.2	
5 - 1	29.9	
6 - 1	28.6	
7 - 1	28.1	Cusum 3S
8 - 1	30.2	
9 - 1	30.5	
11 - 1	29.7	
12 - 1	29.4	
13 - 1	29.8	
14 - 1	28.6	
16 - 1	29	
18 - 1	29.7	
19 - 1	29.2	
20 - 1	29.8	
21 - 1	30.2	
22 - 1	28.6	
26 - 1	28.8	
27 - 1	30.4	
28 - 1	28.5	
29 - 1	29.3	
3 - 1	29	

Westgard	تفسیر	V D3	آزمایش	HPLC Agilent 1100 -1	دستگاه	6	ماه	1402	سال
Control	آزمایشگاه	Mean	29.58	SD	1.1	Exp. Date	1402/6/30	Lot. No	
Next Month	Mean	29.32	SD	0.68	CV	2.31			

