

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal	
Mean	5.99
SD	0.13
CV	2.09
Day-Run	Value
7 - 1	6
3 - 1	5.9
4 - 1	6.1
5 - 1	6
6 - 1	5.9
7 - 2	6.1
8 - 1	6.2
9 - 1	5.9
10 - 1	6.2
12 - 1	6
13 - 1	5.79
13 - 2	6
15 - 1	5.9
16 - 1	6.1
18 - 1	5.9
19 - 1	6.1
20 - 1	5.8
22 - 1	6.1
24 - 1	5.9
25 - 1	6.1
26 - 1	5.9
27 - 1	5.8

High - CV : 1.14	
Mean	15.09
SD	0.17
CV	1.14
Day-Run	Value
7 - 1	15
3 - 1	15.3
4 - 1	15.2
5 - 1	14.9
6 - 1	15.2
7 - 2	15.2
8 - 1	14.9
9 - 1	15
10 - 1	15.3
12 - 1	15.3
13 - 1	15.3
13 - 2	15
15 - 1	15.2
16 - 1	15
18 - 1	15.2
19 - 1	15.1
20 - 1	14.8
22 - 1	15.1
24 - 1	14.9
25 - 1	14.7
26 - 1	15.1
27 - 1	15.2

Westgard	تفسیر	WBC	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	7	ماه	1402	سال
Normal Control	Labex	Mean	6	SD	0.12	Exp. Date	1402/6/30	Lot. No	02-2305n

