

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	6.63
SD	0.12
CV	1.79
Day-Run	Value
1 - 1	6.7
2 - 1	6.5
4 - 1	6.8
5 - 1	6.7
6 - 1	6.6
7 - 1	6.6
8 - 1	6.7
9 - 1	6.5
11 - 1	6.6
12 - 1	6.5
13 - 1	6.8
14 - 1	6.7
18 - 1	6.5
19 - 1	6.5
20 - 1	6.6
21 - 1	6.8
22 - 1	6.5

High - CV : 1.06	
Mean	17.49
SD	0.19
CV	1.06
Day-Run	Value
1 - 1	17.5
2 - 1	17.4
4 - 1	17.7
5 - 1	17.5
6 - 1	17.3
7 - 1	17.9
8 - 1	17.4
9 - 1	17.5
11 - 1	17.5
12 - 1	17.6
13 - 1	17.8
14 - 1	17.2
18 - 1	17.4
19 - 1	17.3
20 - 1	17.5
21 - 1	17.6
22 - 1	17.3

Westgard	تفسیر	WBC	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	6	ماه	1402	سال
Normal Control	Labex	Mean	6.67	SD	0.13	Exp. Date	1402/6/30	Lot. No	02-2305n

