

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	170.53
SD	2.34
CV	1.37
Day-Run	Value
1 - 1	170
2 - 1	172
4 - 1	170
5 - 1	165.5
6 - 1	170
7 - 1	170
8 - 1	175
11 - 1	170
12 - 1	174
13 - 1	168
14 - 1	169
18 - 1	170
19 - 1	173
20 - 1	168
21 - 1	170
22 - 1	170
26 - 1	171
27 - 1	175
28 - 1	170
29 - 1	170

High - CV : 1.65	
Mean	229.68
SD	3.79
CV	1.65
Day-Run	Value
1 - 1	230
2 - 1	233
4 - 1	234
5 - 1	228
6 - 1	227
7 - 1	229
8 - 1	230
11 - 1	230
12 - 1	226
13 - 1	228
14 - 1	231
18 - 1	234
19 - 1	225
20 - 1	235
21 - 1	227
22 - 1	230
26 - 1	238
27 - 1	224
28 - 1	225

Westgard	تفسیر	Triglycerides	آزمایش	اتوآنالایزر BS-800	دستگاه	6	ماه	1402	سال
Normal Control	biorex	Mean	170.41	SD	2.54	Exp. Date	1403/1/17	Lot. No	035

