

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	173.42
SD	3.37
CV	1.94
Day-Run	Value
1 - 1	170
3 - 1	178
4 - 1	176.1
6 - 1	170.6
7 - 1	178
8 - 1	170
10 - 1	175.6
11 - 1	168.9
13 - 1	173
14 - 1	170
15 - 1	175
16 - 1	170
20 - 1	179
21 - 1	170
22 - 1	172.7
23 - 1	170.6
24 - 1	173.6
28 - 1	176.7
29 - 1	178
30 - 1	172.5

High - CV : 1.33	
Mean	531.8
SD	7.09
CV	1.33
Day-Run	Value
1 - 1	530
3 - 1	525
4 - 1	541.2
6 - 1	526.2
7 - 1	539
8 - 1	521
10 - 1	538.4
11 - 1	523.4
13 - 1	532
14 - 1	540
15 - 1	525.7
16 - 1	523.8
20 - 1	540
21 - 1	531.9
22 - 1	538.4
22 - 2	538.4
23 - 1	523.1
24 - 1	542.2
28 - 1	525.2
29 - 1	533
30 - 1	530

Westgard	تفسیر	Ikalan Phosphatase(ALP)	آزمایش	اتوآنالایزر BS-800	دستگاه	8	ماه	1402	سال
Normal Control	biorex	Mean	173	SD	2.6	Exp. Date	1403/1/17	Lot. No	035

