

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.2	
Mean	38.94	Mean	144.32
SD	0.82	SD	1.73
CV	2.12	CV	1.2
Day-Run	Value	Day-Run	Value
1 - 1	38.5	1 - 1	143
3 - 1	38	3 - 1	142.8
4 - 1	37.9	4 - 1	144.8
6 - 1	39.4	6 - 1	145.4
7 - 1	38.9	7 - 1	144.7
8 - 1	39.3	8 - 1	143.8
10 - 1	40.1	10 - 1	141.9
11 - 1	38.8	11 - 1	146.2
13 - 1	39	13 - 1	143
14 - 1	39.6	14 - 1	143
15 - 1	38.1	15 - 1	145
16 - 1	38.5	16 - 1	143
20 - 1	40	20 - 1	148.1
21 - 1	39.1	21 - 1	143
22 - 1	38	22 - 1	145.1
23 - 1	37.2	23 - 1	141.6
24 - 1	39.5	24 - 1	147.1
28 - 1	40.4	28 - 1	146.1
29 - 1	39.4	29 - 1	143.7
30 - 1	39	30 - 1	145

Westgard	تفسیر	SGOT (AST)	آزمایش	اتوآنالایزر BS-800	دستگاه	8	ماه	1402	سال
Normal Control	biorex	Mean	39.21	SD	0.72	Exp. Date	1403/1/17	Lot. No	035

