

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	48	
2 - 1	47.9	
3 - 1	47.1	
4 - 1	47.6	
6 - 1	48.3	
7 - 1	47.5	
8 - 1	47.5	
9 - 1	46.9	
10 - 1	46.3	
11 - 1	47.7	
13 - 1	47.5	
14 - 1	46.8	
15 - 1	47.4	
16 - 1	48.1	
17 - 1	47	
20 - 1	47.2	
22 - 1	46.4	
23 - 1	47.2	
24 - 1	48.1	
25 - 1	47.8	
29 - 1	47.5	
30 - 1	48	

Westgard	تفسیر	CRP-quantitative	آزمایش	BS-480 اتوآنالایزر	دستگاه	8	ماه	1402	سال
Control	biorex	Mean	47.31	SD	0.55	Exp. Date	1403/6/8	Lot. No	326-2-034/3
Next Month	Mean	47.45	SD	0.54	CV	1.14			

