

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	175.35
SD	2.58
CV	1.47
Day-Run	Value
1 - 1	178
3 - 1	178
4 - 1	180
6 - 1	175
7 - 1	174
8 - 1	173
10 - 1	178
11 - 1	176
13 - 1	174
14 - 1	178
15 - 1	174
16 - 1	176
20 - 1	172
21 - 1	179
22 - 1	173
23 - 1	175
24 - 1	173
28 - 1	176
29 - 1	175
30 - 1	170

High - CV : 1	
Mean	264.65
SD	2.64
CV	1
Day-Run	Value
1 - 1	266
3 - 1	263
4 - 1	267
6 - 1	264
7 - 1	264
8 - 1	266
10 - 1	260
11 - 1	269
13 - 1	268
14 - 1	264
15 - 1	262
16 - 1	268
20 - 1	268
21 - 1	262
22 - 1	267
23 - 1	262
24 - 1	262
28 - 1	266
29 - 1	262
30 - 1	263

Westgard	تفسیر	Cholesterol	آزمایش	اتوآنالایزر BS-800	دستگاه	8	ماه	1402	سال
Normal Control	biorex	Mean	175.48	SD	2.29	Exp. Date	1403/1/17	Lot. No	035

